

Contact Person:

Name: Pierrot, Denis
Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory
Address: 4301 Rickenbacker Causeway, Miami Fl, 33149
Phone: 305-361-4441
Email: Denis.Pierrot@noaa.gov

Investigator(s):

Name: Wanninkhof, Rik
Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory
Address: 4301 Rickenbacker Causeway, Miami Fl, 33149
Phone: 305-361-4379
Email: Rik.Wanninkhof@noaa.gov

Name: Pierrot, Denis
Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory
Address: 4301 Rickenbacker Causeway, Miami Fl, 33149
Phone: 305-361-4441
Email: Denis.Pierrot@noaa.gov

Dataset Information:

Funding_Info: NOAA Climate Program Office; NOAA Ocean Acidification Program
Initial_Submission: 2013mmdd
Revised_Submission: 2013mmdd

Cruise Information:

Experiment Name: GU1005_DWH_leg1
Experiment Type: Research Cruise
Platform Type: Ship
Co2 Instrument Type: Equilibrator-IR or CRDS or GC

Cruise ID: 33GG20101008
Cruise Info: AOML_SOOP_CO2
Geographical Region:

Westernmost Longitude: -87.6
Easternmost Longitude: -84.1
Northernmost Latitude: 29.9
Southernmost Latitude: 26.9

Cruise Dates (YYYYMMDD)

Start_Date: 20101016
End_Date: 20101026

Ports of Call:

Pascagoula, MS

Vessel Name: R/V Gordon Gunter
Vessel ID: 33GG
Vessel Owner: NOAA

Variables Information:

Variable Name: xCO2_EQU_ppm

Description of Variable: Mole fraction of CO2 in the equilibrator headspace (dry) at equilibrator temperature (ppm)

Unit of Variable: ppm

Variable Name: xCO2_ATM_ppm

Description of Variable: Mole fraction of CO2 measured in dry outside air (ppm)

Unit of Variable: ppm

Variable Name: xCO2_ATM_interpolated_ppm

Description of Variable: Mole fraction of CO2 in outside air associated with each water analysis. These values are interpolated between the bracketing averaged good xCO2_ATM analyses (ppm)

Unit of Variable: ppm

Variable Name: PRES_EQU_hPa

Description of Variable: Barometric pressure in the equilibrator headspace (hPa)

Unit of Variable: hPa

Variable Name: PRES_ATM@SSP_hPa

Description of Variable: Barometric pressure measured outside, corrected to sea level (hPa)

Unit of Variable: hPa

Variable Name: TEMP_EQU_C

Description of Variable: Water temperature in equilibrator (°C)

Unit of Variable: Degree C

Variable Name: SST_C

Description of Variable: Sea surface temperature (°C)

Unit of Variable: Degree C

Variable Name: SAL_permil

Description of Variable: Sea surface salinity on Practical Salinity Scale (o/oo)

Unit of Variable: ppt

Variable Name: fCO2_SW@SST_uatm

Description of Variable: Fugacity of CO2 in sea water at SST and 100% humidity (µatm)

Unit of Variable: µatm

Variable Name: fCO2_ATM_interpolated_uatm

Description of Variable: Fugacity of CO2 in air corresponding to the interpolated xCO2 at SST and 100% humidity (µatm)

Unit of Variable: µatm

Variable Name: dfCO2_uatm

Description of Variable: Sea water fCO2 minus interpolated air fCO2 (µatm)

Unit of Variable: µatm

Variable Name: WOCE_QC_FLAG

Description of Variable: Quality control flag for fCO2 values (2=good, 3=questionable)

Unit of Variable: None

Variable Name: QC_SUBFLAG

Description of Variable: Quality control subflag for fCO2 values, provides explanation when QC flag=3

Unit of Variable: None

Method Description:

Equilibrator Design:

Depth of Seawater Intake: 5 meters

Location of Seawater Intake: Bow

Equilibrator Type: Spray head above dynamic pool, no thermal jacket

Equilibrator Volume: 0.95 L (0.4 L water, 0.55 L headspace)

Water Flow Rate: 1.5 - 2.0 L/min

Headspace Gas Flow Rate: 70 - 150 ml/min

Vented: Yes

Drying Method for CO₂ in Water:

Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).

Additional Information: Primary equilibrator is vented through a secondary equilibrator.

CO₂ in Marine Air:

Measurement: Yes, 5 readings in a group every 3 hours

Location and Height: Bow mast, ~18 meters above sea surface

Drying Method:

Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).

CO₂ Sensor:

Measurement Method: IR

Manufacturer: LI-COR

Model: 7000

Frequency: Every 140 seconds, except during calibration

Resolution Water: $\pm 0.01 \mu\text{atm}$ in fCO₂_SW

Uncertainty Water: $\pm 2 \mu\text{atm}$ in fCO₂_SW

Resolution Air: $\pm 0.01 \mu\text{atm}$ in fCO₂_ATM

Uncertainty Air: $\pm 0.5 \mu\text{atm}$ in fCO₂_ATM

Manufacturer of Calibration Gas:

Std 1: LL100000, 0.00 ppm, owned by AOML, used every ~2.5 hours. Std 2: JA02280, 248.73 ppm, owned by AOML, used every ~2.5 hours. Std 3: JA02292, 372.88 ppm, owned by AOML, used every ~2.5 hours. Std 4: JA02689, 520.79 ppm, owned by AOML, used every ~2.5 hours.

Number of Non Zero Gas Standards: 2

CO₂ Sensor Calibration:

The analyzer is calibrated every 3 hours with field standards that in turn were calibrated with primary standards that are directly traceable to the WMO scale. The zero gas is ultra-high purity air.

Other Comments:

Instrument is located in an air-conditioned laboratory. Ultra-High Purity air (0.0 ppm CO₂) and the high standard gas are used to zero and span the LI-COR analyzer.

Method References:

Pierrot, D., C. Neil, K. Sullivan, R. Castle, R. Wanninkhof, H. Lueger, T. Johannessen, A. Olsen, R. A. Feely, and C. E. Cosca (2009), Recommendations for autonomous underway pCO₂ measuring systems and data reduction routines, Deep-Sea Res II, 56, 512-522.

Details Co₂ Sensing:

details of CO₂ sensing (not required)

Measured Co₂ Params:

xco2(dry)

Sea Surface Temperature:

Location: hull mounted, ~3 m below sea surface

Manufacturer: Furuno

Model: T2000

Accuracy Degrees Celsius: 0.2

Precision Degrees Celsius: 0.1

Calibration: Factory calibration

Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Equilibrator Temperature:

Location: Inserted into equilibrator ~5 cm below water level

Manufacturer: Hart

Model: 1521

Accuracy Degrees Celsius: 0.025

Precision Degrees Celsius: 0.001

Calibration: Factory calibration

Comments: Resolution is taken as Precision.

Equilibrator Pressure:

Location: Attached to equilibrator headspace. Combined with Licor Pressure

Manufacturer: Licor

Model: None

Accuracy hPa: 1.2

Precision hPa: 0.02

Calibration: Factory calibration

Comments:

Differential pressure reading from Setra-239 attached to the equilibrator headspace was added to the pressure reading from the LICOR analyzer to yield equilibrator pressure. Manufacturer's Resolution is taken as Precision.

Atmospheric Pressure:

Location: Next to the bridge, ~15 m above the sea surface water

Manufacturer: RMYoung

Model: 61201

Accuracy: ± 0.5 hPa

Precision: 0.01 hPa

Calibration: Factory calibration

Normalized: yes

Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Sea Surface Salinity:

Location: In Chem lab, next to CO2 system

Manufacturer: Seabird

Model: SBE 21

Accuracy: ± 0.05 o/oo

Precision: 0.002 o/oo

Calibration: Factory calibration

Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Additional Information:

After Year Day 288, Atmospheric Pressures were estimated by adding 3 mbar to the LICOR P (value obtained

from offset between Licor and atm P from 286 to 288). The pressures associated with ATM measurements have been flagged 3 (questionable P) Issues with Standard tanks not being open for a while: Up to Year Day 289, no tanks were open, data was discarded. From 289 to 296, only 2 tanks were open (Std1 and 2) and data has been flagged 3. Original Data Location: http://www.aoml.noaa.gov/ocd/ocdweb/gunter/gunter_introduction.html

Preliminary Quality Control:

NA

Form Type:

underway